

GOVERNMENT OF JAMMU AND KASHMIR  
DIRECTORATE OF AGRICULTURE TALAB-TILLO, JAMMU

Sub : Modifications and insertions


Ref: This office NIT No. Acctt/PS/NIT/2014-15/1076-1100 dated 10/02/2015

**Corrigendum**

The NIT issued vide this office NIT No. mentioned above , the following below mentioned Modifications and Insertions are needed , which were not incorporated with the already loated NIT. The Modifications and insertions are as below ;

1. Specification for HPLC Systems: As per available literature from NIPHM Hyderabad the following insertions need to be done.  
Since PDA detector is used for residual analysis and to be replaced by UV-Vis Spectrophotometer detector needed for formulation analysis. Besides, injector port should be automatically adjusted so that pre-calibrated sample loop which will allow only fixed amount of sample solution on the column every time
2. Specification for Gas chromatograph : Gas chromatograph should be supplied with FID and TCD detectors required for formulation analysis. Besides, detectors should be sensitive, should have large linear range , should have satiability in single output with time, should show selective response. Besides, GC should be supplied with packed columns either stainless steel or glass having internal diameter of 2-4 mm preferably should comply BIS specifications.

No: Acctt/PS/NIT/2014-15/ 1136-44  
Dated : 23/02/2015

Yours Faithfully  
  
Member secretary (SLRCFC)  
Director of Agriculture  
Jammu

Copy to the :

1. Director of Agriculture Lal Mandi Kashmir (Chairman)
2. Director of Agriculture Talab-Tillo Jammu
3. Director of Finance APD Civil Secretariat Jammu
4. Director Law Enforcement Civil Secretariat Jammu
5. Director Horticulture Jammu
6. Director Horticulture Kashmir
7. Joint Director of Agriculture Inputs Jammu
8. Joint Director of Agriculture Inputs Kashmir .
9. Computer Programmer

for information and n/action